

CERTIFICATE OF CONFORMANCE

 Customer STOCK

A. General Information

Product Type	High Purity Helium	Batch No.	256044
Purity	≥ 99.9995%	Filling Date	January 08, 2026
Cylinder Size	47L	Date of Analysis	January 10, 2026
Cylinder Pressure	150 Bar	Expiry Date	January 08, 2029
Connection Type	CGA718		
Cylinder No.	020071	053081	055733
	058517	12D096107	14D018125
	80018213	80018410	D470292
	D470355	D471008	D472307
	D7052038	M4258193	M4262047
	M5284049		

B. Specification

Analyzed Impurities	Specification
Moisture (Dew Point)	≤ -76.0 °C (1.0 ppm)
Oxygen	≤ 1.0 ppm
THC as Methane	≤ 0.5 ppm
Carbon Dioxide	≤ 0.1 ppm
Carbon Monoxide	≤ 0.1 ppm
Nitrogen	≤ 3.0 ppm



Approved By

14/01/2026

Date



Certified By

QA-FM-35-00

Leeden National Oxygen Ltd

 1 Shipyard Road, Singapore 628128 / Sales Office: 21 Tanjong Kling Road, Singapore 628047
 Tel: (+65) 6265 1222 / (+65) 6265 9355 • Fax: (+65) 6265 1706 • Customer Service Hotline: (+65) 6262 2221
 Email: sgsgas@leedennox.com • Website: apac.nipponsanso.com/leeden-national-oxygen